



Institute of Nano Science and Technology
Habitat Centre, Sector 64, Phase X, Mohali -160062, Punjab, India
Ph: +91-172-2210073/75; Fax: +91-172-2211074; E-mail ID: purchase@inst.ac.in

NOTICE INVITING TENDER

INST invites tender from the reputed manufacturers or their authorized dealers so as to reach this office on or before scheduled date and time for the instrument, as per specifications given in the Annexure attached to the Tender form. All offers should be made in English and should be written in both figures and words. Tender forms can be downloaded from the website (www.inst.ac.in) of the Institute.

The bidders are requested to read the tender document carefully and ensure compliance with all specifications/instructions herein. Non-compliance with specifications/instructions in this document may disqualify the bidders from the tender exercise. The Director, INST reserves the right to select the item (in single or multiple units) or to reject any quotation wholly or partly without assigning any reason. Incomplete tenders, amendments and additions to tender after opening or late tenders are liable to be ignored and rejected.

INSTRUMENTS LIST

S. No.	Ref. No.	Instrument	Qty	Tender Submission Date and Time	Tender Opening Date and Time	Tender Fee (Non-refundable)	EMD
2	INST/12(279)/2016-Pur	Opto-Mechanical Components	As per attached specification	17/04/2017 upto 2.00PM	17/04/2017 at 3.00PM	Rs. 500/-	Rs 9,000/-

Note: The Institute shall not be responsible for any postal delay about non-receipt / non delivery of the bids or due to wrong addressee.

TERMS AND CONDITIONS

Important Conditions of the tender to be abide by the tenderer

- Due date:** The tender has to be submitted before the due date. The offers received after the due date and time will not be considered.
- Preparation of Bids:** The offer/bid should be submitted in single bid system (i.e.) Technical bid cum financial bid. **The Quotations should be valid for 120 days from the due date.** The Quotations duly sealed and super scribed on the envelope with the reference No. and due date, should be addressed to “**The Director, Institute of Nano Science and Technology, Habitat Centre, Sector 64, Phase –X, Mohali, 160062, Punjab**” so as to reach on or before the due date.
- Delivery of the tender:** The tender shall be sent to the addressee given in para 2 (above) either by post or by courier so as to reach our office before the due date specified in our Schedule. The offer/bid can also be dropped in the tender box on or before the due date specified in the schedule. The tender box is kept in Foyer area of INST.
- Opening of the tender:** The offer/bid will be opened by a committee duly constituted for this purpose. The bidders if interested may participate on the tender opening Date and Time. The



bidder should produce authorization letter from their company to participate in the tender opening. Only one representative will be allowed to participate in the tender opening.

5. **Acceptance/Rejection of bids:** The INST reserves the right to reject any or all offers without assigning any reason.
6. **EMD:** Tender fee/EMD is to be obtained from the bidders except those who are registered with the Central Purchase Organisation, National Small Industries Corporation (NSIC) or the concerned Ministry or Department. The tenderer should submit Tender Fee/EMD amount as per tender ref. no. through NEFT/RTGS in INST Account. Account Details are as follows:
 - a. Name of Beneficiary: Institute of Nano Science and Technology (INST)
 - b. Account No. **2452201001102**
 - c. Name of Bank: **Canara Bank, Sector 34, Chandigarh**
 - d. IFS Code: **CNRB0002452**
 - e. MICR Code: **160015003**
 - f. Swift Code: **CNRBINBBFFC**

The details of transaction for EMD viz. Name of bidder firm, Tender Description, Transaction ID/No. of Transfer, Transaction date, Amount of Transaction, Name of Bank, Address of Bank shall be furnished by the tenderer on their letterhead separately along with their tender.

7. **Refund of EMD:** Bid securities of the unsuccessful bidders shall be returned to them at the earliest after expiry of the final bid validity and latest on or before the 30th day after the award of the contract. In case of successful Tenderer, it will be retained till the successful and complete installation of the equipment.
8. **Performance Security:** The supplier shall be required to submit the performance security in the form of irrevocable bank guarantee issued by any Indian Nationalized Bank for an amount which is equal to the 10% of Purchase value at the time of the installation of the equipment covering warranty period of the equipment and should be kept valid for a period of 60 days beyond the date of completion of warranty period.
9. **REASONABILITY OF PRICES :**

Please quote best minimum prices applicable for a premier Research Institution, **leaving no scope for any further negotiations on prices.** **The quoting party should give a certificate to the effect that** the quoted prices are the minimum and they have not quoted the same item on lesser rates than those being offered to INST to any other customer nor they will do so till the validity of offer or execution of the purchase order, whichever is later. We request you to fill the price reasonability certificate format in the enclosed file (Annexure "1")

The party must give details of identical or similar equipment, if any, supplied to any CSIR labs/DBT Institutes during last three years along with the final price paid and Performance certificate from them.
10. **Force Majeure:** The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.



- For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
 - If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
11. **Risk Purchase Clause:** In event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from the other source on the total risk of the supplier under risk purchase clause.
12. **Packing Instructions:** Each package will be marked on three sides with proper paint/indelible ink, the following:
- i. Item Nomenclature
 - ii. Order/Contract No.
 - iii. Country of Origin of Goods
 - iv. Supplier's Name and Address
 - v. Consignee details
 - vi. Packing list reference number
13. **Delivery of Goods:**
Delivery should be given at **Institute of Nano Science and Technology, Habitat Centre, Sector - 64, Phase 10, Mohali** within a maximum of one month's time from the date of placement of purchase order.
14. **Delayed delivery:** If the delivery is not made within the due date for any reason, INST will have the right to impose penalty @ 0.5% per week and the maximum deduction is 10% of the contract value / price.
15. **Prices:** The price should be quoted in net per unit (after breakup) and must include all packing and delivery charges. The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. However the percentage of taxes & duties shall be clearly indicated. The price should be quoted without custom duty and excise duty, since INST is exempted from payment of Excise Duty and is eligible for concessional rate of custom duty. Necessary certificate will be issued on demand. **(Please refer Annexure – 2 for the price to be quoted).**



16. **Notices:** For the purpose of all notices, the following shall be the address of the Purchaser and Supplier.

i. **Purchaser:** The Director,
Institute of Nano Science and Technology,
Habitat Centre, Sector – 64, Phase X, Mohali – 160064.

ii. **Supplier:** (To be filled in by the supplier)

.....
.....
.....

17. **Resolution of Disputes:** The dispute resolution mechanism to be applied pursuant shall be as follows:

i. In case of Dispute or difference arising between the Purchaser and the supplier relating to any matter arising out of or connected with this contract, such disputes or difference shall be settled in accordance with the Indian Arbitration & Conciliation Act, 1996, the rules there under and any statutory modifications or re-enactments thereof shall apply to the arbitration proceedings. The dispute shall be referred to the Director, Institute of Nano Science and Technology (INST) Mohali and if he is unable or unwilling to act, to some other person appointed by him willing to act as such Arbitrator. The award of the arbitrator so appointed shall be final, conclusive and binding on all parties to this order.

18. **Applicable Law:** The place of jurisdiction would be Mohali (Punjab) INDIA.

19. **Right to Use Defective Goods**

If after delivery, acceptance and installation and within the guarantee and warranty period, the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.

20. **Training**

The Supplier is required to train the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.

21. **Installation & Demonstration**

The supplier is required to do the installation and demonstration of the equipment within one month of the arrival of materials at the INST site of installation, otherwise the penalty clause will be the same as per the supply of materials.



22. **Warranty:** Three Years Warranty shall have to be provided by the firm. The Warranty should be comprehensive on site.
23. **Taxes and Duties**
Suppliers shall be entirely responsible for all taxes, duties, license fees, octroi, road permits, etc., incurred until delivery of the contracted Goods to the Purchaser. However, VAT in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order.
24. **Payment:** 100% payment shall be made by the Purchaser after delivery, inspection, successful installation, commissioning and acceptance of the equipment at INST in good condition and to the entire satisfaction of the Purchaser and on production of unconditional performance bank guarantee as specified in Clause 8 of tender terms and conditions.
25. **User list:** Brochure detailing technical specifications and performance, list of industrial and educational establishments where the items enquired have been supplied must be provided.
26. **Manuals and Drawings**
- Before the goods and equipments are taken over by the Purchaser, the Supplier shall supply operation and maintenance manuals. These shall be in such details as will enable the Purchaser to operate, maintain, adjust and repair all parts of the works as stated in the specifications.
 - The Manuals shall be in the ruling language (English) in such form and numbers as stated in the contract.
 - Unless and otherwise agreed, the goods equipment shall not be considered to be completed for the purposes of taking over until such manuals and drawing have been supplied to the Purchaser.
27. **Site Preparation:** The supplier shall inform to the Institute about the site preparation, if any, needed for the installation of equipment, immediately after the receipt of the purchase order. The supplier must provide complete details regarding space and all the other infrastructural requirements needed for the equipment, which the Institute should arrange before the arrival of the equipment to ensure its timely installation and smooth operation thereafter.
- The supplier shall visit the Institute and see the site where the equipment is to be installed and may offer his advice and render assistance to the Institute in the preparation of the site and other pre-installation requirements.
28. **Acknowledgement:** It is hereby acknowledged that we have gone through all the conditions mentioned above and we agree to abide by them.



SIGNATURE OF TENDERER

ALONG WITH SEAL OF THE COMPANY WITH DATE



BID PARTICULARS

1. Name of the Supplier :
2. Address of the Supplier :
3. Availability of demonstration of equipment : Yes / No
4. Tender cost enclosed: : Yes/No if yes
5. Online EMD submission information enclosed : Yes / No if Yes

Transaction ID/No. of Transfer_____

Transaction date: _____

Amount of Transaction_____

Name of Bank_____

Address of Bank_____

6. Name and address of the Officer/contact person to whom all references shall be made regarding this tender enquiry

Name:

Address:

Ph:

Fax:

Mobile:

Email:

Web:



Compliance statement for the tender specifications

Tender Ref No.: _____

S.No.	Check list of documents/Undertakings	Yes/No	Remarks (give explanation if the answer is No)
1.	Is EMD details attached? (if applicable)		
2.	Is the bidder original equipment manufacturer (OEM)/authorised dealer?		
3.	If authorised dealer, recent dated certificate to this effect from OEM, attached or not?		
4.	Undertaking from OEM regarding technical support & extended warranty period		
5.	Validity of 180 days or not?		
6.	Price Reasonability Certificate enclosed as per format??		
7.	Undertaking from bidder regarding acceptance of tender terms & conditions		
8.	Whether list of reputed users (along with telephone numbers of contact persons) for the past three years specific to the instrument attached?		
9.	Does the instrument comply with all the specifications detailed? Attach a separate sheet showing compliance with the specifications and explanations thereto if the equipment varies from the requested specifications.		
10.	Whether free Installation, Commissioning and Application Training offered?		
11.	Whether comprehensive onsite warranty offered?		
12.	Whether Annual maintenance after expiry of comprehensive onsite warranty quoted separately?		
13.	Whether free of cost shifting of instrument from transit campus to main campus offered?		



Annexure – 1

PRICE REASONABILITY CERTIFICATE

This is to certify that we have offered the maximum possible discount to you in our Quotation No. _____ dated _____ for (Value Rs.) _____.

We would like to certify that the quoted price are the minimum and we have not quoted the same item on lesser rates than those being offered to INST to any other customer nor we will do so till the validity of offer or execution of purchase order, whichever is later.

Seal and Signature of the tenderer



Annexure - 2

Name of Equipment with Model No: _____

Name & Address of Manufacturer: _____

Price Quoted

A. For Indigenous Equipment

- I. Cost of equipment (Rs.) _____
- II. Taxes (CST/VAT) (Rs.) _____
- III. Any other charges (Rs.) _____
- IV. Total cost of equipment (in Rs.) F.O.R, INST, Mohali _____

B. For Imported Equipment

- I. Cost of equipment (In Foreign Currency) _____
- II. Packing and Forwarding _____
- III. FOB/FCA Price _____
- IV. Freight charges upto Delhi Airport _____
- V. Insurance (110% of the cost of the equipment) _____
- VI. CIF/CIP upto Delhi Airport _____

Signature of Tenderer with date and seal

TECHNICAL SPECIFICATIONS OF OPTO-MECHANICAL COMPONENTS

S.No.	DETAILS	QTY
1.	Ø12.7 mm Post Holders Spring-Loaded Hex-Locking Thumbscrew L=50 mm	10
2.	Ø12.7 mm Post Holders, Spring-Loaded Hex-Locking Thumbscrew, L=75 mm	5
3.	Ø12.7 mm Optical Post SS M4 Setscrew M6 Tap L = 30 mm	5
4.	Ø12.7 mm Optical Post SS M4 Setscrew M6 Tap L = 50 mm	10
5.	Ø12.7 mm Optical Post SS M4 Setscrew M6 Tap L = 75 mm	5
6.	Ø12.7 mm Optical Post SS M4 Setscrew M6 Tap L = 100 mm	5
7.	Ø12.7 mm Graduated Optical Post SS M4 Setscrew M6 Tap L = 50 mm	1
8.	Ø12.7 mm Graduated Optical Post SS M4 Setscrew M6 Tap L = 75 mm	1
9.	Ø12.7 mm Translating Optical Post SS M4 Setscrew M6 Tap L = 50.8 mm to 57.2 mm	1
10.	Mounting Base 25 mm x 58 mm x 10 mm, 1 slot, Straight, slot length 27.9 mm	5
11.	Mounting Base, 50 mm x 75 mm x 10 mm, 2 slot, Straight, slot length 31.8 mm	5
12.	Mounting Base with Radial Through Slots, 50 mm x 75 mm x 10 mm	1
13.	Dovetail Rail Carrier 1.00" x 1.00" (25.4 mm x 25.4 mm), 1/4" (M6) counter bored hole for mounting , Counter-bore adapter ring should be include	5
14.	Dovetail Optical Rail, 300 mm, Metric, 8-32 (M4 x 0.7) tapped holes on both side end	3
15.	Stage Height Adapter for 62.5 mm (2.46") Deck Height Metric, that can accept standard 1/4"-20 (M6) metric screws	1
16.	N-BK7 Plano-Convex Lens Ø1" f = 25.4 mm Uncoated, Diameter 1", Diopter +39.4, Radius of Curvature(mm) 13.1, Center Thickness(mm) 11.7, Edge Thickness (mm) 1.8, Back Focal Length (mm) 17.6 or better	1
17.	N-BK7 Plano-Convex Lens Ø1" f = 30.0 mm Uncoated, Diameter 1", Diopter +33.3, Radius of Curvature(mm) 15.5, Center Thickness(mm) 8.6, Edge Thickness (mm) 2.0, Back Focal Length (mm) 24.2 or better	2
18.	N-BK7 Plano-Convex Lens Ø1" f = 35.0 mm Uncoated, Diameter 1", Diopter +28.6, Radius of Curvature(mm) 18.0, Center Thickness(mm) 7.2, Edge Thickness (mm) 2.0, Back Focal Length (mm) 30.1 or better	1
19.	N-BK7 Plano-Convex Lens Ø1" f = 60.0 mm Uncoated, Diameter 1", Diopter +16.7, Radius of Curvature(mm) 30.9, Center Thickness(mm) 4.7, Edge Thickness (mm) 2.0, Back Focal Length (mm) 56.7 or better	1
20.	N-BK7 Plano-Convex Lens Ø1" f = 75.0 mm Uncoated, Diameter 1", Diopter +13.3, Radius of Curvature(mm) 38.6, Center Thickness(mm) 4.1, Edge Thickness (mm) 2.0, Back Focal Length (mm) 72.0 or better	2
21.	N-BK7 Plano-Convex Lens Ø1" f = 100.0 mm Uncoated, Diameter 1", Diopter +10, Radius of Curvature(mm) 51.5, Center Thickness(mm) 3.6, Edge Thickness (mm) 2.0, Back Focal Length (mm) 97.3 or better	1
22.	N-BK7 Plano-Convex Lens Ø1" f = 125.0 mm Uncoated, Diameter 1", Diopter +8.0, Radius of Curvature(mm) 64.4, Center Thickness(mm) 3.3, Edge Thickness (mm) 2.0, Back Focal Length (mm) 122.4 or better	1

23.	N-BK7 Plano-Convex Lens Ø1" f = 150.0 mm Uncoated, Diameter 1", Diopter +6.7, Radius of Curvature(mm) 77.3, Center Thickness(mm) 3.1, Edge Thickness (mm) 2.0, Back Focal Length (mm) 147.5 or better	1
24.	N-BK7 Plano-Convex Lens Ø1" f = 200.0 mm Uncoated, Diameter 1", Diopter +5.0, Radius of Curvature(mm) 103.0, Center Thickness(mm) 2.8, Edge Thickness (mm) 2.0, Back Focal Length (mm) 197.5	1
25.	Ø1" Lens Mount with Internal and External SM1 Threads M4 Tap: Ø1" Lens Mount with Internal and External SM1 Threads M4 Tap, Max Optic Thickness 0.28" (7.1 mm or better, Clear Aperture Ø0.90" (Ø23.0 mm) or better, Optical Height 0.87" (22.1 mm) or better	1
26.	Ø1" Lens Mount with SM1 Internal Threads and No Retaining Lip M4 tap: Ø1" Lens Mount with SM1 Internal Threads and No Retaining Lip M4 Tap, Max Optic Thickness 0.24" (6.1 mm) or better, Clear Aperture Ø1.01" (Ø25.6 mm) or better, Optical Height 0.87" (22.1 mm) or better	5
27.	Alignment Plate for Ø1" Fixed Optic Mounts: Alignment Plate with eight concentric circles or better but should fit Ø1" Fixed Optic Mounts and lens tubes	1
28.	SM1 Spanner Wrench, Graduated, Length = 98mm or larger, additionally an engraved depth scale with metric increments 0.5 mm or better	1
29.	Kinematic Mirror Mount for Ø25.4 mm Optics with Post-Centered Front Plate M4 Taps: Kinematic Mirror Mount for Ø25.4 mm Optics with Post-Centered Front Plate M4 Taps, centers the optic 1.25" (31.8 mm) above the post.	2
30.	Kinematic Mirror Mount For Microscope Objectives Kinematic Mirror Mount For Microscope Objectives, For 0.800"-36 RMS Microscope Objective Thread Angular Range: ±4° or better Resolution: 8 mrad (0.5°) per rev or better 1/4"-80 Stainless Steel Lead Screws with Brass Bushings or better Post Mountable via 8 (M4) Counterbored Clearance Hole or better	1
31.	N-BK7 Right-Angle Isosceles Prism Uncoated 25 mm (base), 25 mm(height), unbeveled, precision 90° edge between the two legs, dimensional tolerance: 0.1 mm or better, angular tolerance: 3 arcmin or better	2
32.	Large Adjustable Clamping Arm, M4 Threaded Post , That can holds Optics up to 40 mm Tall or better	1
33.	Ø150 mm Rotating Breadboard, M6 x 1.0 (25 mm pitch), working height is 12.7 mm or better,	1
34.	Ø90 mm or larger, SM1-Threaded Center Panel for the bread board in 33, M6 Taps	1
35.	Lens Mount with Retaining Ring for Ø1" Optics, M4 Tap: Lens Mount with Retaining Ring for Ø1" Optics, M4 Tap, Max Optic Thickness 7.1 mm, Clear Aperture Ø0.90" (Ø23.0 mm), Optical Height 0.87" (22.1 mm) or better	5
36.	Ring-Actuated SM1 Iris Diaphragm: One side of the iris should be externally threaded, while the other should be internally threaded	1
37.	Round Fiber Bundle, Ø1.3 mm, Low OH, SMA to SMA, 1 m Wavelength Range 400 - 2200 nm or better, NA 0.22 or better, Core Diameter 200 ± 8 µm or better, Cladding Diameter 240 ± 5 µm or better, Short-Term Bend Radius 6.5 cm or better, Long-Term Bend Radius 13 cm or better,	1
38.	Rotation Platform: Ø2" Manual Rotation Stage, Metric, The stage should offer 360° of continuous rotation with a	1

	scale that is engraved in 2° (or smaller) intervals along the outer circumference of the rotating platform. It should be marked every 20° (or smaller). Rotation should be lockable manually.	
39.	SM1 Zoom Housing for Ø1" Optics Non-Rotating 4.1 mm Travel (or better) : SM1 Zoom Housing for Ø1" Optics Non-Rotating 4.1 mm Travel, Non-Rotating Translation of SM1 Lens Tubes or Ø1" Optics up to 0.42" (10.7 mm) Thick, 0.16" (4.1 mm) of Travel at 0.5 mm Per Revolution, Fixed External SM1 and Translating Internal SM1 Mounting Threads	1
40.	Ø1" Adjustable Lens Tube, 0.81" Travel Range or more, External Thread length Ø25.4 mm or better	1
41.	SM1 Lens Tube, 3.00" Thread Depth, One Retaining Ring Included, Bore should accept Ø1" or Ø25.4 mm Optics	1
42.	SM1 Lens Tube, 1.00" Thread Depth, One Retaining Ring Included, Bore should accept Ø1" or Ø25.4 mm Optics	1
43.	SM1 Slotted Lens Tube 3.00"; Thread Depth 2 Retaining Rings Included	1
44.	Adapter with External SM1 Threads and Internal M26 x 0.706 Threads, for Ø1" or Ø25.4 mm Optics	1
45.	Adapter with External M26 x 0.706 Threads and Internal SM1 Threads, for Ø1" or Ø25.4 mm Optics	1
46.	Adapter with External C-Mount Threads and External SM1 Threads, for Ø1" or Ø25.4 mm Optics	1
47.	Ø1.20" Clamp for SM1 Lens Tubes, Inner Diameter 1.21" (30.5 mm)	1
48.	Mini Series Rotation Platform, Metric: The stage should be 360° continuous rotatable with a scale of 2° or smaller intervals along the outer circumference of the rotating platform. Furthermore, it should be marked every 20° (or better) mark and a 20 arcmin Vernier scale. Rotation should be lockable by tightening manually.	1
49.	Rotation Mount for Ø1" (25.4 mm) Optics, M4 Tap: Rotation Mount for Ø1" (25.4 mm) Optics, M4 Tap, The aperture is threaded with our internal SM1 threading (1.035"-40) and accepts Ø1" optics up to 0.47" (11.9 mm) thick, A 360° laser-engraved scale marked at 2° increments, A top-located setscrew can be tightened with a 5/64" (2.0 mm) hex key or balldriver to lock the rotation of the mount	1
50.	2" x 2" or larger Linear Polarizer Sheet Wavelength Range 400 - 700 nm or better Extinction Ratioa >100:1 (400 - 500 nm) or better >1000:1 (500 - 700 nm) or better >5000:1 (530 - 690 nm) or better Polarizing Material Dichroic Film Dimensional Tolerance +0/-0.1 mm Thickness 0.3 mm or thinner Acceptance Angle ±30° or better	2
51.	RMS Microscope Objective Mount, Metric, OMR Microscope Objective Mount should be Equipped with 8-32 (M4) or M6 Tap for Post Mounting	1
52.	Adapter with Internal M4 x 0.7 Threads and External M6 x 1.0 Threaded Stud, Adapter with Internal M4 x 0.7 Threads and External M6 x 1.0 Threaded, Length 0.68" (17 mm), Outer Diameter 0.40" (10 mm)	5
53.	Adapter with External M6 x 1.0 Threads and External M4 x 0.7 Threads Adapter with External M6 x 1.0 Threads and External M4 x 0.7 Threads, Length 0.61" (15 mm), Outer Diameter 0.40" (10 mm)	5
54.	Dual Threaded Adapter with Internal M4 x 0.7 Threads and External M6 x 1.0 Threads: Dual Threaded Adapter with Internal M4 x 0.7 Threads and External M6 x 1.0 Threads, Length 0.39" (10 mm)	5

55.	Adapter with Internal M6 x 1.0 Threads and Internal M4 x 0.7 Threads Adapter with Internal M6 x 1.0 Threads and Internal M4 x 0.7 Threads, Length 0.5" (13 mm), Outer Diameter 0.4" (10 mm)	4
56.	Mini-Series Adapter with External M4 Threads and Internal M3 Threads, Length 0.5" (13 mm)	5
57.	Mini-Series Adapter with External M6 Threads and Internal M3 Threads, Length 0.55" (14 mm)	5
58.	SM1 Zoom Housing for Ø1" Optics, Non-Rotating 2" (50.8 mm) Travel, Non-Rotating Translation of SM1 Lens Tubes and Ø1" Optics up to 0.3" (7.6 mm) Thick, Fixed External and Translating Internal SM1 (1.035"-40) Mounting Threads	1
59.	Ø1" N-BK7 Plano-Convex Lens Kit , Uncoated, 14 pc or more Focal Lengths (mm) 25.4, 35, 50, 75, 100, 125, 150, 200, 250, 300, 400, 500, 750, 1000 or more	1

*****All items should be provided by single manufacturer.**

All Items should have metric compatibility.

Each of the above parts should be mountable on each other wherever required.

The item under #59 maybe or may not be procured depending upon the funds availability after the final evaluation.